



An adaptive optics module for deep tissue multiphoton imaging in vivo

Cristina Rodríguez ^{1,2,3,12}, Anderson Chen ^{2,8,12}, José A. Rivera¹, Manuel A. Mohr ⁴, Yajie Liang^{2,9}, Ryan G. Natan ^{2,3}, Wenzhi Sun^{2,10,11}, Daniel E. Milkie ², Thomas G. Bifano^{5,8}, Xiaoke Chen⁴ and Na Ji ^{1,2,3,6,7} ✉

Understanding complex biological systems requires visualizing structures and processes deep within living organisms. We developed a compact adaptive optics module and incorporated it into two- and three-photon fluorescence microscopes, to measure and correct tissue-induced aberrations. We resolved synaptic structures in deep cortical and subcortical areas of the mouse brain, and demonstrated high-resolution imaging of neuronal structures and somatosensory-evoked calcium responses in the mouse spinal cord at great depths in vivo.

Imaging living organisms with subcellular resolution is key for understanding biological systems. Two-photon (2P) fluorescence microscopy is an essential tool for observing cells and biological processes under physiological conditions deep inside tissue. Three-photon (3P) fluorescence microscopy further extends the imaging depth in opaque tissues^{1–3}, as longer excitation wavelengths experience reduced tissue scattering and out-of-focus excitation is suppressed due to the third-order nonlinear optical process. However, living biological tissues aberrate the wavefront of the excitation light for both 2P and 3P microscopy beyond superficial depths, leading to an enlarged excitation focus of diminished focal intensity, ultimately limiting the in vivo imaging performance to cellular resolution at depth.

In the mouse central nervous system, reliably imaging the structure and function of neurons at subcellular resolution requires that the distortion of the excitation light, which accumulates as the light travels through tissues, be measured and corrected by adaptive optics (AO) methods^{4–7}. How aberrations are measured is the main factor distinguishing the different implementations of AO in microscopy. While direct wavefront measurement methods have high measurement speed, they require separate wavefront sensors and correctors. In addition, the introduction of exogenous near-infrared emitting dyes is required to maintain the performance of direct wavefront sensing in highly scattering samples such as tissues^{8,9}. Albeit slower, indirect wavefront sensing methods are well suited for opaque media. Approaches based on modal wavefront sensing introduce aberration modes extending over the entire excitation wavefront (for example, Zernike polynomials) and evaluate the impact of the introduced aberration on selected image metrics to obtain the corrective wavefront. In contrast, zonal methods segment the wavefront into zones and measure the wavefront error in each zone. Using two wavefront modulation devices, a zonal indirect wavefront sensing method based on frequency-multiplexed aberration

measurement achieved 2P imaging of submicrometer-sized dendritic spines in layer 5 of the mouse primary visual cortex¹⁰. However, this method was limited to a narrow band of wavelengths, low power throughput and slow measurement times by the liquid-crystal spatial light modulator used for aberration correction. Here, we report a compact AO module for multiphoton microscopy, composed of a high-speed segmented deformable mirror (DM), two lenses and a field stop, that overcomes the aforementioned limitations. We demonstrate high-resolution 2P and 3P fluorescence imaging in a variety of optically challenging locations in the mouse central nervous system. With AO, we were able to resolve fine neuronal processes and synapses (for example, dendritic spines) in deep cortical layers as well as subcortical areas of the mouse brain in vivo: features otherwise invisible without aberration correction. In the mouse spinal cord, we achieved high-resolution in vivo imaging of neuronal structures at depths beyond 400 μm and somatosensory-evoked calcium responses at depths surpassing 300 μm .

Results

A compact AO module for fluorescence microscopy. To build a compact AO module, we implemented the frequency-multiplexed aberration measurement¹⁰ and the subsequent correction solely by the DM (Extended Data Fig. 1). We first determined the phase gradient values to be added to each DM segment such that the beamlets reflecting off them would converge to a single point on the focal plane. To achieve this, we formed a stationary reference focus using half of the beamlets and raster-scanned the other half of the beamlets around this focus by tip-tilting the corresponding DM segments. At each set of phase gradients (produced by tipping and tilting the mirror segments leading to beamlet displacements on the focal plane), we recorded the fluorescence signal variation while modulating the phase or intensity of the scanned beamlets at distinct roughly kHz frequencies using the DM segments.

¹Department of Physics, University of California, Berkeley, CA, USA. ²Janelia Research Campus, Howard Hughes Medical Institute, Ashburn, VA, USA.

³Department of Molecular and Cell Biology, University of California, Berkeley, CA, USA. ⁴Department of Biology, Stanford University, Stanford, CA, USA.

⁵Department of Mechanical Engineering, Photonics Center, Boston University, Boston, MA, USA. ⁶Helen Wills Neuroscience Institute, University of California, Berkeley, CA, USA. ⁷Molecular Biophysics and Integrated Bioimaging Division, Lawrence Berkeley National Laboratory, Berkeley, CA, USA.

⁸Present address: Bio Optical and Acoustic Spectroscopy Laboratory, Neurophotonics Center, Boston University, Boston, MA, USA. ⁹Present address: Department of Diagnostic Radiology and Nuclear Medicine, University of Maryland School of Medicine, Baltimore, MD, USA. ¹⁰Present address: School of Basic Medical Sciences, Capital Medical University, Beijing, China. ¹¹Present address: Chinese Institute for Brain Research, Beijing, China. ¹²These authors contributed equally: Cristina Rodríguez, Anderson Chen. ✉e-mail: jina@berkeley.edu

We obtained the interference strength of each beamlet with the reference focus by Fourier transforming the signal trace and recording the Fourier magnitude at its modulation frequency. We then calculated the phase gradients giving rise to maximal interference strength for these beamlets. By swapping the scanned and the stationary DM segments and repeating the phase gradient measurement, we obtained the wavefront gradients required to converge all beamlets to the same focus. We then measured the phase of each beamlet directly following a similar modulation-based procedure¹¹, again taking advantage of the high-speed DM segments. For larger aberrations, the whole procedure may be repeated for several iterations to achieve optimal aberration correction. We then applied the final corrective wavefront to the DM for aberration correction. After reflecting off the DM and traveling through the sample, all beamlets converged around a common spot with the same phase so that they constructively interfered to form a diffraction-limited focus. In contrast to previous work¹⁰, where a slow DM or digital micromirror device was used for frequency-multiplexed modulation, and a spatial light modulator for aberration measurement and correction, using a single high-speed DM both simplifies our module and enables higher power throughput (up to 15 \times), along with polarization- and wavelength-independent operation. These advantages allowed us to use the same module for 2P and 3P fluorescence microscopy under a wide range of excitation wavelengths (for example, 900–1,700 nm), and enabled aberration measurement and correction at great depths. Moreover, compared to previous implementations¹⁰, our module enables faster (roughly 4–20 \times) aberration measurement, which is particularly important in the context of 3PM, where the longer excitation wavelengths generate more heating due to higher tissue absorption.

We incorporated our AO module into a homebuilt 2P fluorescence microscope by placing it between the excitation laser and the microscope (Extended Data Fig. 2). We first validated its performance in correcting artificial aberrations using the signal from fluorescent features of different sizes. For one representative artificial aberration (Extended Data Fig. 3), both intensity and phase modulation enabled corrections that substantially increased signal and improved resolution of fluorescent bead images, with higher signal recovery for smaller beads (23.7 \times for 0.5- μ m diameter beads versus 3.7 \times for 10- μ m diameter beads).

Adaptive optical 2P fluorescence imaging in vivo. We next applied the AO module to high-resolution in vivo imaging. We imaged green fluorescent protein (GFP)-labeled myotomes in the mid-trunk of a young zebrafish larva, using a 920-nm laser excitation wavelength (Fig. 1a–c). At an imaging depth of 110 μ m below the surface, the aberration introduced by this sample, mostly astigmatism resulting from the larva's highly curved cylindrical-like surface, led to images of low resolution and contrast (Fig. 1a). Aberration correction resulted in an increase of the 2P fluorescence signal of roughly 1.2–5.1 \times (Supplementary Fig. 1a–d) and a substantial improvement in image contrast (roughly 2–4.5 \times , Supplementary Fig. 1a–d) and resolution (spatial frequency space representation, insets in Fig. 1a).

We also achieved substantial improvements in image quality for in vivo 2P imaging of the strongly scattering mouse brain through a cranial window (Fig. 1d–i). At imaging depths of 370 and 500 μ m below dura, aberration correction improved the signal of yellow fluorescent protein (YFP)-labeled dendrites by up to roughly 2.7 \times and the image contrast by 2–12 \times (Supplementary Fig. 1e–h), enabling the identification of fine features such as dendritic spines and axonal boutons, which in some cases can only be resolved in the aberration-corrected image (Fig. 1g,h). In the spatial frequency space (insets in Fig. 1d,g and Extended Data Fig. 4), the resolution improvement gained through aberration correction led to a substantial increase in the magnitudes of high spatial frequency components. Here, and in the following examples, the noncorrected

(‘No AO’) images were taken after correcting aberrations intrinsic to the microscope system (Extended Data Fig. 2) and adjusting the objective correction collar to remove the spherical aberration added by the coverglass used in sample preparation. These ‘No AO’ images, therefore, represent the best performance that conventional optics and best practice can achieve. The improvements in signal, resolution and contrast therefore only arose from AO correction of tissue-induced aberrations.

Adaptive optical 3P fluorescence imaging in the mouse brain in vivo. To push the imaging depth even further, we incorporated our AO module into a homebuilt 3P microscope (Extended Data Fig. 2). We validated the module's ability to recover nearly diffraction-limited imaging performance under large wavefront aberrations and observed a 270 \times signal improvement for beads in a capillary tube (Extended Data Fig. 5).

We then used our AO 3P fluorescence microscope to image neuronal structures throughout the mouse cerebral cortex in vivo. As an example, at 1,300 nm excitation, a single correction pattern applied to the DM improved the image quality of YFP-labeled neurons (Fig. 2a–e, Extended Data Fig. 6 and Supplementary Videos 1 and 2), located 760 μ m below dura. Here, AO correction led to 3P signal increases ranging from six to eight times on the cell body, to eight to 30 times on dendritic structures (Fig. 2c and Supplementary Fig. 2), with larger signal increases observed in finer features. Moreover, previously invisible synaptic structures (for example, dendritic spines and spine necks) became easily detectable only after aberration correction (Fig. 2b). In all mouse brain examples shown here, we used only two to three rounds of correction to obtain the final corrective wavefront. Additional rounds did not result in substantial improvement of the fluorescence signal (Extended Data Fig. 7). The data shown are representative of several imaging sessions performed at similar depths (600–870 μ m below dura), in different regions of the same animal, as well as in several animals ($N=8$) and under different labeling conditions (Extended Data Figs. 6 and 8 and Supplementary Videos 3 and 4). Aberration correction consistently increased the image signal and contrast throughout fields of view of up to 200 μ m (Supplementary Fig. 3), which is consistent with the isoplanatic patch size of 100–200 μ m measured previously using 2P microscopy^{8,10,12}. Moreover, the improvement in spatial resolution gained through aberration correction—evidenced by the recovery of spatial frequency components as shown in the spatial frequency space representations of the images (Fig. 2e and Extended Data Figs. 6 and 8)—allowed us to resolve synaptic structures throughout cortex down to 870 μ m below dura (Extended Data Fig. 8), which were otherwise not identifiable without correction.

Our AO module enabled imaging of hippocampal structures with subcellular resolution through the highly scattering white matter, without the need to perform any invasive procedure (for example, cortical tissue removal, optical device insertion) other than the cranial window implantation (Fig. 2f–m and Supplementary Videos 5–8). Using 1,300 nm excitation, aberration correction improved the 3P fluorescence signal of YFP-labeled neuronal structures—4 \times on the cell body (Fig. 2f and Supplementary Video 5) and 2–7 \times on dendritic structures (Fig. 2g and Supplementary Fig. 2)—and the recovery of high spatial frequency components (Fig. 2j), allowing synaptic features (for example, dendritic spines) to be clearly resolved only after AO correction (Fig. 2g and Supplementary Video 7). In addition to 3P excitation of green fluorophores using 1,300 nm, excitation of red fluorophores with 1,700 nm has been shown to improve tissue penetration¹. To assess the performance of our module under 1,700 nm excitation, we imaged tdTomato-labeled neuronal cell bodies at 952–1,020 μ m below dura, and found a similar improvement in the image signal and contrast after AO correction (Fig. 2k–m, Extended Data




Fig. 1 | AO improves in vivo 2P imaging of myotomes in zebrafish larva and neuronal structures in the mouse brain. **a**, Lateral and axial (along red dashed line) images of myotomes in the mid-trunk of a 4-day-old zebrafish larva Tg(β -actin:HRAS-EGFP), at an imaging depth of 110 μm from the surface, without and with AO. Postobjective power: 13 mW. **b**, Signal profiles in the axial plane along the yellow lines in **a**. **c**, Corrective wavefront in **a**. **d–i**, Imaging of dendrites in the cerebral cortex of Thy1-YFP-H mice. **d**, Maximum intensity projections of dendrites at 365–375 μm below dura imaged without and with AO. **e**, Signal profiles along to the purple lines in **d**. **f**, Correction wavefront in **d**. **g**, Maximal intensity projection of dendrites at 490–513 μm below dura imaged without and with AO. **h**, Signal profiles along the blue lines in **g**. **i**, Corrective wavefront in **g**. Phase modulation. Excitation wavelength: 920 nm. Postobjective power: 31 mW (**d**) and 128 mW (**g**). Insets in **a**, **d** and **g** show the spatial frequency space representations of the corresponding fluorescence images, dashed circles show diffraction-limited resolution. Scale bars, 10 μm in **a** and 5 μm in **d** and **g**. Microscope objective NA 0.8 16 \times for **a** and NA 1.05 25 \times for **d** and **g**. Zebrafish imaging shows representative results from three fields of view and two zebrafish larvae, and mouse brain imaging shows representative results from 13 fields of view and three mice. AU, arbitrary units.

Fig. 9 and Supplementary Video 8). Emerging brighter fluorescent sensors would allow us to further push the depth limits for which an accurate aberration measurement can be obtained and substantial improvement in image quality can be achieved. Last, the large increase in 3P image signal and contrast observed in the mouse brain after correcting for tissue-induced aberrations enabled us to limit the postobjective average power to 10–40 mW (see Supplementary Table 1 for all imaging parameters).

Adaptive optical 3P fluorescence imaging in the mouse spinal cord in vivo. The imaging performance attained by incorporating our AO module into a 3P microscope motivated us to image neuronal structures in an even more challenging environment of the central nervous system: the spinal cord. The high neuronal density¹³, along with the surface curvature and strong scattering caused by superficial axon tracts, cause wavefront distortions that have prevented 2P microscopy from imaging neuronal structures beyond the most superficial <200 μm of the spinal cord dorsal horn^{14–16}, which in its entirety spans a depth of roughly 500 μm . We performed in vivo 3P imaging of GFP-labeled neuronal structures at depths exceeding 400 μm below dura through a dorsal laminectomy¹³, in 9- to 10-week-old adult mice, under 1,300 nm excitation (Fig. 3a–g). After aberration correction, we achieved an improvement in signal ranging from 2–7.1 \times (Fig. 3c,f and Supplementary

Fig. 4) and resolution (insets in Fig. 3b), which allowed us to more clearly visualize fine neuronal structures.

Finally, our AO module enabled us to reliably record somatosensory-evoked calcium transients in the spinal cord dorsal horn at depths beyond 300 μm . Due to its optically challenging nature, such recordings had been limited to the most superficial (<100 μm) layers^{13,14,17}, preventing a comprehensive understanding of the complex coding of somatosensory stimuli in the spinal cord circuitry. We recorded calcium transients in jGCaMP7-expressing spinal cord neurons¹⁸, in response to cooling applied to the hindlimb skin, at imaging depths where temperature responses had not been previously studied (Fig. 3h–j and Supplementary Video 9)¹³. At an imaging depth of 310 μm below dura, aberration correction substantially improved the signal of a neuronal cell body (by 5.9 \times) and increased the calcium transient amplitudes, with a 2.1 \times increase in the peak calcium-dependent fluorescence change ($\Delta F/F_0$) (Fig. 3j). Although imaging depths exceeding 400 μm were possible, we did not find cooling-responsive neurons at such depths.

Comparison of modal and zonal approaches. To compare modal and zonal representations of aberrations, we simulated the performance of zonal and modal wavefront corrections (Supplementary Fig. 5 and Supplementary Table 2). Assuming perfect performance of modal methods without the practical limitations imposed by




Fig. 2 | AO enables in vivo 3P imaging of cortical and hippocampal neuronal structures in the mouse brain, with subcellular resolution. **a**, Maximum intensity projection of a neuron in the mouse cortex (Thy1-YFP-H), at 747–767 μm below dura, under 1,300 nm excitation, without and with AO (phase modulation). Postobjective power: 17 mW. **b**, Zoomed-in views of the red square in **a**, at 751–767 μm below dura, without and with AO. Insets show zoomed-in views of the dendrite in white rectangles in **b**. 10 \times digital gain was applied to No AO inset to increase visibility. Postobjective power: 20 mW. **c**, Signal profiles along the purple and blue lines in **a**. **d**, Corrective wavefront in **a** and **b**. **e**, Spatial frequency space representations of the images in **b** (left) and their radially averaged profiles (right). **f**, Lateral and axial images of neurons in the mouse hippocampus (Thy1-YFP-H), 719 μm below dura, under 1,300 nm excitation, without and with AO (phase modulation). Postobjective power: 16 mW. Insets show zoomed-in views of the gray square in **e**. The 7 \times digital gain was applied to increase visibility. **g**, Maximum intensity projection of neuronal processes above the cell body in **e**, at 695–709 μm below dura, without and with AO. White arrows, dendritic spines. Postobjective power: 26 mW. Three times digital gain was applied to image without AO to improve visibility. **h**, Signal profiles along the blue lines in **e**. **i**, Corrective wavefront in **e** and **f**. **j**, Spatial frequency space representations of the images in **g** (left) and their radially averaged profiles (right). **k**, Lateral and axial images of neurons in the mouse hippocampus (Gad2-IRES-Cre \times Ai14 (Rosa26-CAG-LSL-tdTomato)), 952 μm below dura, under 1,700 nm excitation, without and with AO (phase modulation). Postobjective power: 30 mW. **l**, Signal profiles along the green and yellow lines in **i**. **m**, Corrective wavefront in **k**. Scale bars, 10 μm . Microscope objective NA 1.05 25 \times . Cortical imaging shows representative results from 32 fields of view and eight mice; hippocampal imaging shows representative results from 15 fields of view and four mice.

factors such as signal-to-noise ratio and metrics' sensitivity to small aberrations, for smooth wavefront aberrations, our zonal approach yields similar residual wavefront errors as modal correction using $\lesssim 30$ Zernike modes (as typically done experimentally^{19–21}). While typical fluorescence acquisition times used in this work (23.6 s) are in line with those from modal approaches (roughly 20–40 s)^{20–22}, our acquisition time can be considerably shortened (for example, by 5 \times) without compromising performance (Methods, Supplementary Fig. 6, and Supplementary Table 3). For more complex aberrations, however, we found that correcting for $\gtrsim 50$ Zernike modes would be needed for modal representation to yield similar wavefront errors as our zonal approach, further slowing down modal routines. Therefore, our simulation indicated that zonal methods such as ours, when compared with modal methods, are superior for

complex aberrations encountered at larger tissue depths, in agreement with earlier work comparing modal and zonal representations of aberrations^{23,24}.

Assessment of tissue injury after running AO routine. We investigated whether carrying out aberration measurement with our AO module substantially increased the possibility of tissue damage. Following previously used protocols^{25,26}, we immunolabeled postmortem mouse brain and spinal cord slices against early markers of tissue injury (heat shock protein and microglial activation) after running our AO routine (Supplementary Figs. 7 and 8 and Supplementary Note 4). Our aberration measurement procedure did not lead to any noticeable tissue injury, as indicated by the lack of immunoreactivity. This is likely because our postobjective average




Fig. 3 | AO improves in vivo 3P structural and functional imaging in the mouse spinal cord. **a**, Schematic of in vivo imaging in the dorsal horn of the mouse spinal cord. **b**, Maximum intensity projection of spinal cord neurons (Thy1-GFP-M), 208–228 μm below dura, under 1,300 nm excitation, without and with AO (phase modulation). Insets show spatial frequency space representations of the corresponding fluorescence images. Postobjective power: 18.3 mW. **c**, Signal profiles along the purple lines in **b**. **d**, Corrective wavefront in **b**. **e**, Lateral and axial images of a neuron (Thy1-GFP-M), 414 μm below dura, under 1,300 nm excitation, without and with AO (phase modulation). Postobjective power, 89 mW. **f**, Signal profiles along the blue and yellow lines in **e**. **g**, Corrective wavefront in **e**. **h**, Schematic for recording calcium activity in jGCaMP7s-expressing neurons of the dorsal horn in the mouse spinal cord (AAV8-Syn-jGCaMP7s), in response to cooling stimuli applied to the skin of the hindlimb. **i**, Lateral image of a neuron, 310 μm below dura, under 1,300 nm excitation, after AO correction. **j**, 3P fluorescence signal (top) and calcium transients ($\Delta F/F_0$) (bottom), during temperature stimulation (bottom), without and with AO (phase modulation), for the neuronal cell body shown in **i**. Four-trial average; shaded area, s.e.m. Postobjective power: 4.2 mW. **k**, Corrective wavefront in **i**. Scale bars, 10 μm . Microscope objective NA 1.05 25 \times . Structural imaging shows representative results from seven fields of view and two mice; functional imaging shows representative results from three fields of view and two mice.

laser powers remained well below the levels where heating-related effects^{25,26} or photodamage²⁷ would be a concern.

Discussion

In this work, we combined an add-on AO module based on a zonal aberration measurement method into a 2P and a 3P fluorescence microscope. We achieved substantial improvements in image quality, along with subcellular resolution, on a variety of biological structures in vivo at large depths.

Besides the ability to achieve subcellular resolution at depth, by improving the focus quality through aberration correction, the power delivered to the sample can be substantially reduced while maintaining signal brightness (for example, by up to 1.6 \times and 3.1 \times based on the observed 2P and 3P fluorescence signal improvement, respectively), minimizing the out-of-focus background and thus extending the imaging depth limit.

The high-power throughput, ease of implementation and small footprint of our AO module, along with its polarization- and

wavelength-independent operation, provides easy integration into existing laser-scanning multiphoton microscopes, such as 2P and 3P microscopes, as well as other point-scanning modalities including those based on harmonic generation and Raman scattering. With a standalone software control module, our module can be adopted by a variety of biological laboratories, enabling the investigation of biological processes deep inside living tissues with sub-cellular resolution, in fields ranging from neurobiology and cancer biology to plant biology.

Online content

Any methods, additional references, Nature Research reporting summaries, source data, extended data, supplementary information, acknowledgements, peer review information; details of author contributions and competing interests; and statements of data and code availability are available at <https://doi.org/10.1038/s41592-021-01279-0>.

Received: 25 November 2020; Accepted: 23 August 2021;
Published online: 4 October 2021

References

- Horton, N. G. et al. In vivo three-photon microscopy of subcortical structures within an intact mouse brain. *Nat. Photonics* **7**, 205–209 (2013).
- Ouzounov, D. G. et al. In vivo three-photon imaging of activity of GCaMP6-labeled neurons deep in intact mouse brain. *Nat. Methods* **14**, 388–390 (2017).
- Wang, T. & Xu, C. Three-photon neuronal imaging in deep mouse brain. *Optica* **7**, 947–960 (2020).
- Kubby, J. A. *Adaptive Optics for Biological Imaging* (CRC Press, 2013).
- Booth, M. J. Adaptive optical microscopy: the ongoing quest for a perfect image. *Light.: Sci. Appl.* **3**, e165 (2014).
- Ji, N. Adaptive optical fluorescence microscopy. *Nat. Methods* **14**, 374–380 (2017).
- Rodríguez, C. & Ji, N. Adaptive optical microscopy for neurobiology. *Curr. Opin. Neurobiol.* **50**, 83–91 (2018).
- Wang, K. et al. Direct wavefront sensing for high-resolution in vivo imaging in scattering tissue. *Nat. Commun.* **6**, 7276 (2015).
- Liu, R., Li, Z., Marvin, J. S. & Kleinfeld, D. Direct wavefront sensing enables functional imaging of infragranular axons and spines. *Nat. Methods* **16**, 615–618 (2019).
- Wang, C. et al. Multiplexed aberration measurement for deep tissue imaging in vivo. *Nat. Methods* **11**, 1037–1040 (2014).
- Liu, R., Milkie, D. E., Kerlin, A., MacLennan, B. & Ji, N. Direct phase measurement in zonal wavefront reconstruction using multidither coherent optical adaptive technique. *Opt. Express* **22**, 1619–1628 (2014).
- Ji, N., Sato, T. R. & Betzig, E. Characterization and adaptive optical correction of aberrations during in vivo imaging in the mouse cortex. *Proc. Natl Acad. Sci. USA* **109**, 22–27 (2012).
- Ran, C., Hoon, M. A. & Chen, X. The coding of cutaneous temperature in the spinal cord. *Nat. Neurosci.* **19**, 1201–1209 (2016).
- Johannssen, H. C. & Helmchen, F. In vivo Ca²⁺ imaging of dorsal horn neuronal populations in mouse spinal cord. *J. Physiol.* **588**, 3397–3402 (2010).
- Matsumura, S., Taniguchi, W., Nishida, K., Nakatsuka, T. & Ito, S. In vivo two-photon imaging of structural dynamics in the spinal dorsal horn in an inflammatory pain model. *Eur. J. Neurosci.* **41**, 989–997 (2015).
- Cheng, Y.-T., Lett, K. M. & Schaffer, C. B. Surgical preparations, labeling strategies, and optical techniques for cell-resolved, in vivo imaging in the mouse spinal cord. *Exp. Neurol.* **318**, 192–204 (2019).
- Sekiguchi, K. J. et al. Imaging large-scale cellular activity in spinal cord of freely behaving mice. *Nat. Commun.* **7**, 11450 (2016).
- Dana, H. et al. High-performance calcium sensors for imaging activity in neuronal populations and microcompartments. *Nat. Methods* **16**, 649–657 (2019).
- Débarre, D. et al. Image-based adaptive optics for two-photon microscopy. *Opt. Lett.* **34**, 2495–2497 (2009).
- Galwaduge, P. T., Kim, S. H., Grosberg, L. E. & Hillman, E. M. C. Simple wavefront correction framework for two-photon microscopy of in-vivo brain. *Biomed. Opt. Express* **6**, 2997 (2015).
- Streich, L. et al. High-resolution structural and functional deep brain imaging using adaptive optics three-photon microscopy. Preprint at *bioRxiv* <https://doi.org/10.1101/2021.01.12.426323> (2021).
- Tao, X. et al. Transcutaneous imaging with cellular and subcellular resolution. *Biomed. Opt. Express* **8**, 1277 (2017).
- Panagopoulou, S. I. & Neal, D. R. Zonal matrix iterative method for wavefront reconstruction from gradient measurements. *J. Refract. Surg.* **21**, 563–569 (2005).
- Hu, Q. et al. A universal framework for microscope sensorless adaptive optics: generalized aberration representations. *APL Photonics* **5**, 100801 (2020).
- Podgorski, K. & Ranganathan, G. Brain heating induced by near-infrared lasers during multiphoton microscopy. *J. Neurophysiol.* **116**, 1012–1023 (2016).
- Wang, T. et al. Quantitative analysis of 1300-nm three-photon calcium imaging in the mouse brain. *eLife* **9**, e53205 (2020).
- Olivié, G. et al. Wavelength dependence of femtosecond laser ablation threshold of corneal stroma. *Opt. Express* **16**, 4121–4129 (2008).

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

© The Author(s), under exclusive licence to Springer Nature America, Inc. 2021

Methods

Animals. All animal experiments were conducted according to the National Institutes of Health guidelines for animal research. Procedures and protocols on mice and zebrafish were approved by the Institutional Animal Care and Use Committee at Janelia Research Campus, Howard Hughes Medical Institute; and the Animal Care and Use Committee at the University of California, Berkeley. Mice were housed in cages in groups of 1–5, under a normal light cycle, at a temperature of about 22.8 °C (73 °F) and humidity of 37%. Both males and females were used in this study, with ages ranging from 5 to >30 weeks old and included (Jackson Laboratories) Thy1-YFP-H, Thy1-GFP-M, Gad2-IRES-Cre X Ai14 and C57BL/6J. Details on animal preparations are available below.

Excitation source. The 2P excitation source was a mode-locked titanium:sapphire laser (Chameleon Ultra II, Coherent) operating at 920 nm. The 3P excitation source consisted of a two-stage optical parametric amplifier (Opera-F, Coherent) pumped by a 40-W diode-pumped femtosecond laser (Monaco 1035-40-40, Coherent) operating at 1,035 nm and 1 MHz, providing a broad tuning range (650–920 and 1,200–2,500 nm). Opera-F was operated at 1,300 and 1,700 nm for 3P excitation, for which the average output power was roughly 1.5 and 0.9 W (1.5 and 0.9 μJ per pulse at 1 MHz repetition rate), respectively. For 1,300 nm excitation, to reduce the group delay dispersion (GDD) at the sample plane, we used a homebuilt single-prism compressor²⁸. After compensation, the pulse duration at the focal plane of the objective was measured to be roughly 54 fs using an autocorrelator (Carpe, APE GmbH). For 1,700 nm excitation, since the GDD is anomalous for many of the glasses and crystals used in our microscope, the resulting negative GDD at the sample plane cannot be compensated for by using our prism-based compressor. Instead, the high normal dispersion of ZnSe (from a bulk compressor available inside Opera-F) and silicon (from a 3-mm thick window placed at Brewster's angle²⁹) was used to obtain a pulse duration at the sample of roughly 70 fs after compensation.

Adaptive optical microscope setup. Simplified diagrams of our homebuilt 2P microscope and 3P microscope are shown in Extended Data Fig. 2.

For the 2P microscope (Extended Data Fig. 2a), a pair of achromat doublets (AC254-200-B and AC254-300-B, Thorlabs) conjugated the segmented DM surface to a liquid-crystal spatial light modulator (Holoeye, PLUTO-NIR) used to introduce artificial aberrations. The spatial light modulator plane was then conjugated to a pair of galvanometers (6215H, Cambridge Technology) that were optically conjugated to each other and the back focal plane of a high numerical aperture (NA) water-dipping objective (Olympus XLPLN25XWMP2, NA 1.05, 25× for mouse brain imaging; or Nikon CFI LWD, NA 0.8, 16× for zebrafish imaging), using three pairs of achromat doublets (AC254-150-B and AC254-60-B, AC508-080-B and AC508-080-B, AC508-75-C and SLB-50-600PIR1; Thorlabs and OptoSigma).

For the 3P microscope (Extended Data Fig. 2b), femtosecond pulses at 1,300 or 1,700 nm were reflected off a segmented DM. The DM was conjugated to a pair of galvanometers (6215H, Cambridge Technology) that were optically conjugate to each other and the back focal plane of a high-NA water-dipping objective (Olympus XLPLN25XWMP2, NA 1.05, 25×), using three pairs of achromat doublets (Original 3P system AC254-100-C and 45-804, AC508-080-C and AC508-080-C, AC508-100-C and SLB-50-600PIR2; Thorlabs, Edmund Optics and OptoSigma. New 3P system AC254-400-C and AC254-300-C, SL50-3P and SL50-3P, SL50-3P and TTL200MP; Thorlabs). Here, the DM image underfilled the back aperture of the objective, leading to an effective NA of roughly 0.9, corresponding to a lateral and axial resolutions of roughly 0.6 and 2.3 μm, respectively, at 1,300 nm.

The DM in this project is a hexagonal tip-tilt-piston DM (Hex-111-X) manufactured by the Boston Micromachines Corporation (Supplementary Fig. 9 and Supplementary Note 1). For both the 2P and 3P microscopes, a field stop (iris diaphragm, Thorlabs) was located at the intermediate image plane between the DM and the X galvo to block unwanted diffraction orders and light reflected off mirror segments at large tilt angles. To translate the focus axially, the objective was mounted on a piezoelectric stage (P-725.4CD PIFOC, Physik Instrumente). The fluorescence signal was collected by the same objective and reflected from a dichroic beam splitter (FF665-Di02-25x36, Semrock), spectrally filtered (2P FF01-680/SP, Semrock; 3P FF01-680/SP, Semrock; together with BLP01-442R-25, Semrock, or ET575lp, Chroma), to block the third-harmonic generation signal at 1,300 and 1,700 nm, respectively) and detected by a photomultiplier tube (H7422-40 or H10770PA-40, Hamamatsu). A Pockels cell was used for controlling the excitation power (2P M350-80, 3P M360-40; Cooptics). For the experiments using 1,700 nm excitation, we used D₂O, instead of H₂O, for the objective immersion media, because of its much lower absorption²⁹. Custom-written software was used for image acquisition.

Aberration measurement method. The general aberration measurement procedure was conceptually similar to that described in our previous work¹⁰. Experimentally, the laser focus was parked at one sample location and the fluorescence signal from this location was used for aberration measurement. The pupil was segmented to 37 regions, corresponding to the number of segments

in the DM. The 37 segments were separated into two groups of alternating rows (Extended Data Fig. 1).

First, we fixed the tip, tilt and piston of one group of 17 segments, but added to each of the other 20 pupil segments a specific tip angle θ_i and tilt angle Φ_j ($i, j = 1, 2, \dots, n$) (Extended Data Fig. 1, step 1), each of which were chosen randomly from an array of n angles evenly spaced between $-\Psi/2$ and $\Psi/2$. These applied tip and tilt angles caused the beamlet reflecting off the segment to displace along the x and y axes in the objective focal plane by X_i and Y_j ($X_i = f \times \tan(2\theta_i/M)$ and $Y_j = f \times \tan(2\Phi_j/M)$, where f is the focal length of the objective and M is the magnification from the DM to objective back focal plane), respectively, which changed their interference with the reference focus formed by the other 17 beamlets. With the tip and tilt angles of all segments fixed, using the segments themselves we then modulated the phase or intensity of all 20 beamlets (Supplementary Note 2), each at a distinct frequency ω_s ($s = 1, 2, \dots, 20$), and recorded the fluorescence signal for a time duration T (Extended Data Fig. 1, step 2). The recorded signal trace was then Fourier transformed and the Fourier magnitudes at each distinct modulation frequency ω_s were measured (Extended Data Fig. 1, step 3), whose values indicated how much individual beamlets interfered with the reference focus at the focal displacement of (X_s, Y_s).

The above procedure was repeated $n \times n$ times, so that all 20 segments sampled the full tip and tilt angles and their corresponding beamlets scanned over a two-dimensional (2D) grid in the focal plane with the dimensions $2f \times \tan(\Psi/M)$ by $2f \times \tan(\Psi/M)$ (Extended Data Fig. 1, step 4). For each beamlet, plotting the Fourier magnitudes versus the displacements (X_s, Y_s), we constructed a 2D map of interference strength of this beamlet with the reference focus at different focal displacements (Extended Data Fig. 1, step 5). Fitting the map with a 2D Gaussian function, we found the displacements leading to maximal interference between the beamlets and their reference focus, which gave us the tip and tilt angles (that is, phase gradient) to be applied to this segment in the corrective wavefront.

We then repeated steps 1–5, but now with the group of 20 pupil segments fixed and forming the reference focus. Modulating the remaining 17 segments, we obtained their 2D maps of interference strength versus displacement (Extended Data Fig. 1, step 6) and the phase gradients required to shift the corresponding beamlets to coincide with the reference focus. The total fluorescence signal acquisition time was $2 \times n \times n \times T$.

With all the beamlets intersecting at the same location, the next step was to determine the phase offsets that would allow them to constructively interfere³⁰. Consider the case of finding the phase offset that enables two beamlets to constructively interfere. With one beamlet assigned as the reference (with unknown phase θ_r), by incrementally adding a phase offset $\Delta\theta$ to the other beamlet (with unknown phase θ_i) at step size ω (that is, $\Delta\theta = \omega t$), the intensity variation can be described by $I = 2 + 2\cos(\omega t + \theta_i - \theta_r)$. The phase offset that gives the maximal intensity (that is, constructive interference) is thus the opposite of the phase of the function $\cos(\omega t + \theta_i - \theta_r)$. One approach is by Fourier transforming the time-dependent signal and reading out the phase at frequency $\omega/2\pi$. To determine the phase offsets of 37 segments, we used the concept of multidither coherent optical adaptive technique^{11,31–33}. We modulated the phases of the first 20 rays by piston-displacing each corresponding mirror segment at a distinct frequency ω_s while keeping the phases of the remaining rays, which formed a reference focus, constant (Supplementary Note 3). We then Fourier transformed the recorded fluorescence trace (Extended Data Fig. 1, step 7), and read out the phase offsets that would lead to constructive interference with the reference focus at the modulation frequencies ω_s ($s = 1, 2, \dots, 20$) (Extended Data Fig. 1, step 8). We next modulated the phases of the remaining 17 segments while keeping the phases of the first 20 segments unchanged and found the phase offsets for these 17 segments required to ensure constructive interference among beamlets.

With all beamlets intersecting and constructively interfering at the focal plane, we obtained the final corrective wavefront and applied it to the DM for aberration correction. Because the reference foci used for both phase gradient measurements (Extended Data Fig. 1, steps 1–6) and phase offset measurements (Extended Data Fig. 1, steps 7–9) were aberrated to begin with, for larger aberrations, the whole procedure was iterated, as needed, to achieve optimal aberration correction (Extended Data Fig. 1, step 10).

Typical operation parameters and imaging considerations. An example set of operation parameters used for aberration measurement in the mouse brain in vivo (Fig. 2) included an integration time of $T = 90$ ms for each of the 11×11 ($n \times n$) tip and tilt angles, which scanned the modulated rays over $19 \times 19 \mu\text{m}^2$ 2D grids in the focal plane. The overall fluorescence acquisition time was therefore 21.8 s ($2 \times n \times n \times T$). Additionally, for the phase measurement portion of the algorithm, the total fluorescence acquisition time was 1.8 s (360 ms per iteration, with a total of five iterations). Additional hardware (for example, DM settling time) and software overheads added to the fluorescence acquisition time and the overall time used for aberration measurement and correction was three to four times the fluorescence acquisition time. For the data in Fig. 2, three rounds of aberration measurements were performed to obtain the final corrective wavefront. However, our method does allow for considerably shorter fluorescence acquisition times for the gradient measurement step without affecting performance, because the process of determining the location of centroids in the gradient measurement

step is robust against noise. For cranial window induced aberrations, reducing the integration time for each tip and tilt angle by $8.3\times$ (from 90 to 10.8 ms) and the total sample illumination time by $5.3\times$ from the typical parameters used in this manuscript yielded the same correction performance (Supplementary Fig. 6 and Supplementary Table 3).

'No AO' images were taken after system aberration correction (Extended Data Fig. 2c–e), as well as the adjustment of the objective correction collar to compensate for the spherical aberration introduced by the glass windows overlaying the mouse brain and spinal cord. Furthermore, to minimize additional aberration modes that arose from a tilted window and sample (for example, coma)³⁴, we used the third-harmonic generation signal from the window–tissue interface to determine the angle of the window (Supplementary Fig. 10) and adjusted the mouse using a 2D goniometer stage till the window was perpendicular to the excitation beam. These procedures constitute the best practice and lead to the best performance that conventional optics can achieve. The deterioration in image signal and contrast observed in our 'No AO' images, therefore, resulted from tissue-induced aberrations exclusively and could only be compensated by AO.

Digital image processing. Due to brain motion at depth, the StackReg³⁵ image registration plug-in in ImageJ was used for rigid registration in 2D. The 'smooth' function from ImageJ that replaced the value of each pixel with the average of 3×3 pixels centering on this pixel was applied to all 3P images. The 2P images were presented using the 'Green hot' lookup table in ImageJ and 3P images were presented with the 'Green hot' and 'Magenta hot' lookup tables for 1,300 and 1,700 nm excitation, respectively. For images where the signal was too weak before AO correction, a linear scaling factor was applied to all pixel values to improve visibility with the scaling factor listed on the image and figure legend. The images presented did not undergo any other digital manipulation. For calculating the spatial frequency space representations of fluorescence images, a Gaussian Blur filter (ImageJ) with blur radius $\sigma = 1$ pixel was first applied to the corresponding fluorescence image to eliminate pixelation artifacts that would otherwise show up as high spatial frequency components of nonnegligible amplitude (sometimes past the diffraction limit). The spatial frequency space representations of fluorescence images are shown in logarithmic scale.

Bead samples. Carboxylate-modified fluorescent microspheres (Fluosphere, Invitrogen) were immobilized on poly(L-lysine)-coated microscope slides (12-550-12, Fisher Scientific).

Zebrafish preparation. Zebrafish procedures have been described previously³⁶. Briefly, Tg(β -actin:HRAS-EGFP) zebrafish embryos (*Danio rerio*) were grown at 28 °C in E3 zebrafish embryo medium. To generate optically transparent embryos, melanin synthesis was inhibited by transferring the embryos into 1 \times phenylthiourea (PTU) solution in E3 medium, 10–16 h postfertilization. Before imaging, the chorions were manually removed with forceps under a stereomicroscope. Four-day-old larvae were then anesthetized in E3 medium containing 1 \times tricaine and immobilized on a dish by embedding them in 0.5% low-melting point agarose with 1 \times PTU and 1 \times tricaine. During imaging, E3 medium containing 1 \times PTU and 1 \times tricaine was used as immersion medium.

Mouse preparation (brain imaging). Cranial window implantation procedures were performed³⁷, using aseptic technique, on mice that were anesthetized with isoflurane (1–2% by volume in O₂) and given the analgesic buprenorphine (SC, 0.3 mg per kg of body weight). For 2P imaging experiments, a 3.5-mm diameter craniotomy was made over V1 with dura left intact. A glass window made of two coverslips (Fisher Scientific, thickness no. 1.5) bonded with ultraviolet cured optical adhesives (Norland Optical Adhesives 61) was embedded in the craniotomy and sealed in place with dental acrylic. For 3P imaging experiments, a 5-mm diameter craniotomy was made over V1, with dura left intact. The glass window consisted of a donut-shaped coverslip (inner diameter 4.5 mm, outer diameter 5.5 mm; Potomac Photonics) bonded with ultraviolet cured optical adhesives (Norland Optical Adhesives 61) on top of a 5-mm diameter coverslip (Denville Scientific, thickness no. 1). The window was embedded in the craniotomy and sealed in place with dental acrylic. A titanium head post was attached to the skull with cyanoacrylate glue and dental acrylic. Acute imaging was performed 1 h after the surgery; chronic imaging happened at least 1 week after the surgery. Mice were head-fixed and anesthetized using isoflurane (1–2% by volume in O₂) during imaging.

Some mice also underwent virus injection procedures as described previously³⁷. For the examples shown in Extended Data Fig. 9, neurons in the hippocampus were infected with a mixture of AAV-Syn-Cre ($10\times$ dilution from 1.8×10^{13} GC per ml) and AAV-CAG-FLEX-tdTomato (3.3×10^{13} GC per ml) in a wild-type mouse (C57BL/6J). Three injection sites were chosen (AP -1.7 mm, ML $+1.5$ mm; AP -2.0 mm, ML $+2.0$ mm; AP -2.3 mm, ML $+2.5$ mm), at five different depths (0.6, 0.8, 1, 1.2 and 1.4 μ m). Next, 100 nl of viral solution were injected at each spot. A cranial window was implanted, as described above, 14 d after virus injection and acute imaging was then performed.

Mouse preparation (spinal cord imaging). Acute spinal cord windows were prepared as described previously¹³. Briefly, mice were anesthetized with two

successive intraperitoneal injections of 1 mg kg⁻¹ body weight urethane each, 30 min apart. A tracheotomy was performed and mice were intubated to prevent asphyxiation. The T11–13 vertebrae were exposed and stabilized using spinal clamps (STS-A, Narishige). A dorsal laminectomy was performed at T12 to expose the spinal cord. After a wash with Ringer solution (135 mM NaCl, 5.4 mM KCl, 5 mM HEPES, 1.8 mM CaCl₂, pH 7.2), the spinal cord was covered with a glass window made of a single coverslip (Fisher Scientific No. 1.5). The window and the surrounding custom imaging chamber were stabilized using 2% agarose in Ringer solution. Next, 20 ml kg⁻¹ body weight of 0.9% physiological saline was administered subcutaneously after the surgery, to ensure hydration during the imaging session. Blood flow through the central blood vessel was continuously monitored throughout the imaging experiment to ensure tissue health. For the functional imaging experiments, wild-type (C57BL/6) mice had previously been intrathecally injected with roughly 5×10^{10} GC of AAV8-Syn-jGCaMP7s. Animals were euthanized at the end of the imaging session.

Temperature stimulation (spinal cord imaging). The left hindlimb of the mouse was dehaired and gently fixed inside a custom-designed stimulation device as previously described¹³. The entire limb was continuously exposed to a homogeneous high flow-rate water flow of variable temperature. For each trial, the spinal cord was imaged for a total of 70 s. The first 10 s were imaged at 31 °C baseline water to obtain the baseline fluorescence and noise. For the following 30 s, the flow was switched to water that was preincubated at colder temperatures with the same flow rate. For the remaining 30 s the flow was switched back to the 31 °C baseline water. At least another 120 s passed before the next trial was performed. The actual temperature inside the stimulation chamber was monitored and recorded using a microprobe thermometer (BAT-12, Physitemp) with a Type-K thermocouple placed right next to the mouse's limb, simultaneously with the fluorescence images. The electric valves controlling the water flow switch were triggered by and synchronized with the image acquisition system.

Analysis of calcium imaging data. To account for motion in the spinal cord caused by respiration, we processed the image sequences using an iterative cross-correlation-based registration algorithm³⁷. We averaged across trials, manually selected the region of interest, and calculated the mean fluorescence within this region. From this, we calculated the fractional change in 3P fluorescence ($\Delta F/F_0$) due to neural activity, with F_0 being the baseline fluorescence calculated as the mean fluorescence during exposure to the baseline temperature (initial 10 s of the temperature stimulation). For the traces shown in Fig. 3j, we performed a five-frame moving average.

Reporting Summary. Further information on research design is available in the Nature Research Reporting Summary linked to this article.

Data availability

The main data supporting the findings of this study are available within the paper and its Supplementary information files. The source data files for all data presented in this paper can be found at: <https://github.com/JiLabUCBerkeley/JiLabAO>.

Code availability

A LabVIEW executable program that allows standalone operation of the AO module and detailed documentations of its operation are available for download at <https://github.com/JiLabUCBerkeley/JiLabAO>.

References

- Akturk, S., Gu, X., Kimmel, M. & Trebino, R. Extremely simple single-prism ultrashort-pulse compressor. *Opt. Express* **14**, 10101–10108 (2006).
- Horton, N. G. & Xu, C. Dispersion compensation in three-photon fluorescence microscopy at 1,700 nm. *Biomed. Opt. Express* **6**, 1392–1397 (2015).
- Ji, N., Milkie, D. E. & Betzig, E. Adaptive optics via pupil segmentation for high-resolution imaging in biological tissues. *Nat. Methods* **7**, 141–147 (2010).
- Bridges, W. B. et al. Coherent optical adaptive techniques. *Appl. Opt.* **13**, 291–300 (1974).
- O'Meara, T. R. Theory of multidither adaptive optical systems operating with zonal control of deformable mirrors. *J. Opt. Soc. Am.* **67**, 318–325 (1977).
- O'Meara, T. R. The multidither principle in adaptive optics. *J. Opt. Soc. Am.* **67**, 306–315 (1977).
- Turcotte, R., Liang, Y. & Ji, N. Adaptive optical versus spherical aberration corrections for in vivo brain imaging. *Biomed. Opt. Express* **8**, 3891–3902 (2017).
- Thévenaz, P., Ruttimann, U. E. & Unser, M. A pyramid approach to subpixel registration based on intensity. *IEEE Trans. Image Process.* **7**, 27–41 (1998).
- Godinho, L. Imaging zebrafish development. *Cold Spring Harb. Protoc.* **2011**, 879–883 (2011).
- Sun, W., Tan, Z., Mensh, B. D. & Ji, N. Thalamus provides layer 4 of primary visual cortex with orientation- and direction-tuned inputs. *Nat. Neurosci.* **19**, 308–315 (2016).

Acknowledgements

We thank J. Wu for help with the detection system; the Janelia JET team for designing and assembling the dispersion compensation unit; E. Carroll for help with galvo electronics; S. Chen for surgical assistance and K. Borges and Q. Zhang for helpful discussions. This work was supported by the Howard Hughes Medical Institute (C.R., A.C., Y.L., R.G.N., W.S., D.E.M. and N.J.); the National Institutes of Health U01NS118300 (C.R. and N.J.); the Burroughs Wellcome Fund under the Career Awards at the Scientific Interface (C.R.); Lawrence Berkeley National Laboratory LDRD 20-116 (J.A.R.) and the Firmenich Next Generation Fund, the Terman Fellowship, NIH grant nos. R01DA045664, R01MH116904 and R01HL150566 (X.C.).

Author contributions

N.J. conceived and supervised the project. D.E.M., A.C. and N.J. developed the AO control program. C.R. and A.C. developed the AO standalone control program. T.G.B. performed the DM calibration and provided support on its operation. Y.L., R.G.N. and W.S. performed the mouse brain surgery. M.A.M. performed the mouse spinal cord surgery. M.A.M. built spinal cord temperature stimulation device and C.R. developed control program. A.C. built 2P AO setup and collected 2P data. C.R. built 3P AO setup and collected 3P data in brain and beads. J.A.R. designed new 3P system with input from C.R. J.A.R. and C.R. built new 3P system (used for spinal cord experiments). C.R. and M.A.M. designed and performed 3P imaging experiments in spinal cord. C.R. analyzed data and prepared all figures and supplementary material. X.C. supervised spinal cord

experiments. C.R. and N.J. wrote the manuscript with feedback from M.A.M. and input from all other authors.

Competing interests

N.J. and Howard Hughes Medical Institute have filed patent applications that relate to the principle of frequency-multiplexed aberration measurement. T.G.B. has a financial interest in Boston Micromachines Corporation, which produced commercially the DB used in this work. The remaining authors declare no competing interests.

Additional information

Extended data are available for this paper at <https://doi.org/10.1038/s41592-021-01279-0>.

Supplementary information The online version contains supplementary material available at <https://doi.org/10.1038/s41592-021-01279-0>.


Correspondence and requests for materials should be addressed to Na Ji.

Peer review information *Nature Methods* thanks the anonymous reviewers for their contribution to the peer review of this work. Nina Vogt was the primary editor on this article and managed its editorial process and peer review in collaboration with the rest of the editorial team.


Reprints and permissions information is available at www.nature.com/reprints.

Phase gradient measurement



(1) Apply tip and tilt to first group of segments

(2) Modulate each segment at a distinct frequency ω_s & record signal


(3) FT the signal trace and read the FT magnitudes at the modulation frequencies (orange data points)

(4) Repeat $n \times n$ times for all tip and tilt angles


(5) Plot the FT magnitude for all tip and tilt combinations

Pixel intensity = Fourier magnitude for displacement $[f \cdot \tan(2\Theta/M), f \cdot \tan(2\Phi/M)]$


(6) Repeat (1)-(5) for the second group of segments



Phase offset measurement


(7) Modulate the phase of the first group of segments at a distinct frequency ω_s & record signal

(8) FT the signal trace and read the phases at the modulation frequencies (orange data points)




(9) Repeat (7)-(8) for the second group of segments


(10) Final corrective wavefront




Extended Data Fig. 1 | Schematics of the aberration measurement method. (1) We fix the tip, tilt, and piston of one group of 17 segments, and add to the remaining 20 pupil segments a specific tip angle Θ_i and tilt angle Φ_j ($i, j = 1, 2, \dots, n$) chosen randomly from an array of n angles spaced between $-\Psi/2$ and $\Psi/2$. (2) We modulate the phase or intensity of all 20 beamlets, at a distinct frequency ω_s ($s = 1, 2, \dots, 20$), and record the fluorescence signal. (3) We Fourier transform (FT) the recorded signal trace and measure the Fourier magnitudes at each ω_s . (4) This procedure is repeated $n \times n$ times for all tip and tilt angles. (5) For each beamlet, we plot the Fourier magnitudes versus the displacements (X_i, Y_j) and construct a 2D map of interference strength of this beamlet with the reference focus at different focal displacements. Here, $X_i = f \cdot \tan(2\Theta_i/M)$ and $Y_j = f \cdot \tan(2\Phi_j/M)$; f : focal length of the objective; M : magnification from the DM to objective back focal plane. (6) We repeat steps (1)-(5), with the group of 20 pupil segments fixed, while modulating the remaining 17 segments. We fit each map with a 2D Gaussian function and find the displacements leading to maximal interference between the light ray and the reference focus, corresponding to the tip and tilt angles for each segment of the corrective wavefront. (7) We modulate the phases of the first 20 rays by piston-displacing each corresponding mirror segment at a distinct frequency ω_s while keeping the phases of the remaining rays constant. The resulting fluorescence trace is recorded. (8) We FT the recorded fluorescence trace and read out the phase offsets that would lead to constructive interference with the reference focus at the modulation frequencies ω_s . (9) We modulate the phases of the remaining 17 segments while keeping the phases of the first 20 segments unchanged and obtain the corrective wavefront. (10) We repeat the steps above as needed to obtain the final corrective wavefront and apply it to the DM. See detailed description in Methods.




Extended Data Fig. 2 | Schematics of AO 2P and 3P fluorescence microscopes, and example system correction. **a,b**, Components of AO 2P and 3P fluorescence microscopes, respectively. DM, deformable mirror; SLM, spatial light modulator (used to introduce artificial aberration); L, lenses; X and Y, galvanometers; PMT, photomultiplier tube. **c**, Lateral and axial 3P images of a 1- μm -diameter red fluorescent bead, under 1300 nm excitation, taken without and with AO. Post-objective power: 0.13 mW. **d**, Signal profiles along the purple and yellow lines in **c**. **e**, Corrective wavefront applied to the DM. Scale bar, 1 μm . Microscope objective: NA 1.05 25 \times . Data representative of $n > 20$ bead images taken during $n > 20$ imaging sessions.




Extended Data Fig. 3 | Correcting artificial aberrations for 2P microscopy using phase versus intensity modulation and signal from fluorescent features of different sizes. **a,b**, Artificial aberration introduced with the SLM and corrective wavefront on the DM, respectively. **c,d**, Axial images of 0.5- μ m- and 10- μ m-diameter fluorescent beads, respectively, without AO, with AO (phase modulation), and under ideal aberration-free conditions (without artificial aberrations applied to the SLM). Digital gains were applied to No AO images to increase visibility. Post-objective power: 6 and 2.3 mW, respectively. **e,f**, 2P signal versus iteration number for 0.5- μ m- and 10- μ m-diameter beads, respectively, using phase and intensity modulation. Scale bars: 1 μ m in **c** and 5 μ m in **d**. Microscope objective used: NA 0.8 16 \times . $n=1$ bead in **c** and **d**, respectively.




Extended Data Fig. 4 | AO recovers spatial frequency components in 2P images of neuronal structures in the living mouse brain. a-d, Imaging of dendrites in the cerebral cortex of Thy1-YFP-H mice. **a,c,** Maximum intensity projections of dendrites at 365–375 μm and 490–513 μm below dura, respectively, under 920 nm excitation, without and with AO (same images as in Fig. 1). **b,d,** Spatial frequency space representations of the 2P images in **a** and **c**, respectively (left), and their radially averaged profiles (right). Scale bars, 5 μm . Microscope objective: NA 1.05 25x. Representative results from 13 fields of view and 3 mice.




Extended Data Fig. 5 | AO improves 3P imaging of beads in a capillary tube. **a**, Schematics of sample geometry of 1- μm -diameter fluorescent beads in an air-filled capillary tube. **b**, Lateral and axial (along red dotted line) images of beads without and with AO (phase modulation). Post-objective power: 0.13 mW. Digital gains were applied to No AO images to increase visibility. **c,d**, 3P signal improvement (AO/No AO) and axial full width at half maximum (FWHM) of a representative bead (white arrowhead in **b**) as a function of the iteration #, respectively. **e**, Corrective wavefront, unwrapped (modulo- 2π) for visualization purposes. Scale bar, 5 μm . Microscope objective: NA 1.05 25x. Representative results from 2 imaging sessions.




Extended Data Fig. 6 | AO improves *in vivo* 3P imaging of cortical neurons in the mouse brain. **a**, Lateral and axial images of a neuronal cell body (Thy1-YFP-H), at 757 μm below dura, under 1300 nm excitation, without and with AO (same cell body as in Fig. 2a). Post-objective power: 17 mW. **b**, Signal profiles along the purple and yellow lines in **a**. **c**, Maximum intensity projection (MIP) of same neuron as in **a**, 747–767 μm below dura, under 1300 nm excitation, without and with AO. Post-objective power: 17 mW. **d**, MIP of the yellow square in **c**, at 747–757 μm below dura, without and with AO. Insets in **d**, zoomed-in views of dendrite in white box. 10 \times digital gain was applied to the inset without AO to improve visibility. Post-objective power: 20 mW. **e**, Lateral and axial images of a neuron in the mouse cortex (Thy1-GFP-M), at 687 μm below dura, under 1300 nm excitation, taken without and with AO. Post-objective power: 35 mW. **f**, Signal profiles along the purple and yellow lines in **e**. **g**, Corrective wavefront in **e**. **h**, MIP of a neuron in the mouse cortex (Thy1-GFP-M, different animal than in **e**), at 624–644 μm below dura, under 1300 nm excitation, without and with AO. Post-objective power: 13 mW. **i**, Signal profiles along the purple and blue lines in **h**. **j**, Corrective wavefront in **h**. Insets in **a** and **e**: spatial frequency space representation of the corresponding fluorescence images. Scale bars, 10 μm . Microscope objective: NA 1.05 25 \times . Representative results from 32 fields of view and 8 mice.



Extended Data Fig. 7 | Effect of iterations on 3P fluorescence signal improvement for phase and intensity modulation-based aberration correction in the mouse brain *in vivo*. **a-f**, 3P images of a neuron in the mouse cortex (Thy1-YFP-H), $623 \mu\text{m}$ below dura, under 1300 nm excitation, without AO correction and after running aberration measurement a total of $N=1-5$ iterations. **a,d**, Lateral and axial images of the neuron using phase and intensity modulation, respectively. Post-objective power: 20.8 (**a**) and 23.6 mW (**d**). **b, e**, 3P signal improvement (AO/No AO) with iterations, for phase and intensity modulation, respectively. The plotted signal is the average pixel intensity within a 16×16 -pixel area around the image maximum. **c,f**, Corrective wavefronts measured with phase and amplitude modulation, respectively. Scale bars, $10 \mu\text{m}$. Microscope objective: NA 1.05 25x. Representative results from 3 fields of view and 2 mice.



Extended Data Fig. 8 | AO enables *in vivo* 3P imaging of dendritic spines and axonal boutons in deep layers of the mouse cortex. **a**, Maximum intensity projection (MIP) of a neuron in the mouse cortex (Thy1-YFP-H), at 601–616 μm below dura, under 1300 nm excitation, without and with AO. Post-objective power: 17 mW. **b**, Signal profiles along the purple and blue lines in **a**. **c**, Corrective wavefront in **a**. **d**, MIP of the orange box in **a**, at 609–619 μm below dura, without and with AO. Post-objective power: 25.6 mW. **e**, Spatial frequency space representations of the images in **d** (top) and their radially averaged profiles (bottom). **f, g**, Zoomed-in views of the red and gray boxes in **d**, respectively. 4x digital gain was applied to images without AO to improve visibility. White arrowheads: dendritic spines; orange arrowheads: axonal boutons. **h**, MIP of a neuron in the mouse cortex (Thy1-YFP-H), at 863–875 μm below dura, under 1300 nm excitation, taken without and with AO. Post-objective power: 42 mW. **i**, Zoomed-in views of the red box in **h**. 3x digital gain was applied to the image taken without AO to improve visibility. **j**, Signal profiles along the purple and blue lines in **h**. **k**, Corrective wavefront in **h**. **l**, Spatial frequency space representation of the images in **h** (left) and their radially averaged profiles (right). Scale bars: 10 μm in **a**, **d**, **h**, and **i**; 2 μm in **f** and **g**. Microscope objective: NA 1.05 25x. Representative results from 20 fields of view and 5 mice.



Extended Data Fig. 9 | AO improves *in vivo* 3P imaging of hippocampal structures at different depths in the mouse brain, with 1700 nm excitation. a–l, 3P images of neurons in the mouse hippocampus at different depths. **a,d,g,j**, Lateral and axial images of neurons without and with AO, at 917, 960, 1010, and 1020 μm below dura, respectively. Post-objective powers: 26.5 (**a**), 10 (**d**), 27 (**g**), and 24 mW (**j**). **b,e,h,k**, Signal profiles along the green and yellow lines in **a, d, g**, and **j**, respectively. **c,f,i,l**, Corrective wavefront in **a, d, g**, and **j**, respectively. For **a**, a Gad2-IRES-Cre \times Ai14 (Rosa26-CAG-LSL-tdTomato) mouse was used; for **d, g**, and **j**, neurons in wildtype mice were infected by a mix of AAV-Syn-Cre and AAV-CAG-FLEX-tdTomato. Scale bar, 10 μm . Microscope objective: NA 1.05 25 \times . Representative results from 9 fields of view and 2 mice.

Reporting Summary

Nature Research wishes to improve the reproducibility of the work that we publish. This form provides structure for consistency and transparency in reporting. For further information on Nature Research policies, see our [Editorial Policies](#) and the [Editorial Policy Checklist](#).

Statistics

For all statistical analyses, confirm that the following items are present in the figure legend, table legend, main text, or Methods section.

n/a Confirmed

- The exact sample size (n) for each experimental group/condition, given as a discrete number and unit of measurement
- A statement on whether measurements were taken from distinct samples or whether the same sample was measured repeatedly
- The statistical test(s) used AND whether they are one- or two-sided
Only common tests should be described solely by name; describe more complex techniques in the Methods section.
- A description of all covariates tested
- A description of any assumptions or corrections, such as tests of normality and adjustment for multiple comparisons
- A full description of the statistical parameters including central tendency (e.g. means) or other basic estimates (e.g. regression coefficient) AND variation (e.g. standard deviation) or associated estimates of uncertainty (e.g. confidence intervals)
- For null hypothesis testing, the test statistic (e.g. F , t , r) with confidence intervals, effect sizes, degrees of freedom and P value noted
Give P values as exact values whenever suitable.
- For Bayesian analysis, information on the choice of priors and Markov chain Monte Carlo settings
- For hierarchical and complex designs, identification of the appropriate level for tests and full reporting of outcomes
- Estimates of effect sizes (e.g. Cohen's d , Pearson's r), indicating how they were calculated

Our web collection on [statistics for biologists](#) contains articles on many of the points above.

Software and code

Policy information about [availability of computer code](#)

Data collection

LabVIEW custom-written software was used for image acquisition, for running the AO control program, and for controlling the spinal cord temperature stimulation hardware and recording the temperature. A standalone software module and accompanying documentation for our AO method can be found at: <https://github.com/JiLabUCBerkeley/JiLabAO>. LabVIEW 2013 (64-bit) was used for acquiring the data shown in this article. LabVIEW 2020 (64-bit) was used for developing the standalone software module.

Data analysis

ImageJ 1.52k was used for performing the maximum intensity projection of image stacks, taking line profiles, smoothing images using the "smooth" function (which replaced the value of each pixel with the average of 3×3 pixels centering on this pixel), for registration of structural imaging stacks using the plugin StackReg (latest version from 16/09/2010), and for manually selecting regions of interest in the calcium imaging datasets and calculating the mean fluorescence within this region. MATLAB 2018b was used for making all plots shown in this manuscript, for analyzing the images taken with and without adaptive optics in order to obtain the signal ratio (AO/No AO) and the axial extent of the imaged beads (where applicable), and for analyzing the calcium imaging data (registering the image sequences using an iterative cross-correlation-based algorithm, averaging across trials, calculating the standard error of the mean, calculating the fractional change in fluorescence).

For manuscripts utilizing custom algorithms or software that are central to the research but not yet described in published literature, software must be made available to editors and reviewers. We strongly encourage code deposition in a community repository (e.g. GitHub). See the Nature Research [guidelines for submitting code & software](#) for further information.

Data

Policy information about [availability of data](#)

All manuscripts must include a [data availability statement](#). This statement should provide the following information, where applicable:

- Accession codes, unique identifiers, or web links for publicly available datasets
- A list of figures that have associated raw data
- A description of any restrictions on data availability

The authors declare that the main data supporting the findings of this study are available within the paper, its extended data, and supplementary information files. The Source Data files for all data presented within the Figures and Extended Data can be found at: <https://github.com/JiLabUCBerkeley/JiLabAO> (under the "ManuscriptSourceDataFiles" folder).

Field-specific reporting

Please select the one below that is the best fit for your research. If you are not sure, read the appropriate sections before making your selection.

- Life sciences Behavioural & social sciences Ecological, evolutionary & environmental sciences

For a reference copy of the document with all sections, see [nature.com/documents/nr-reporting-summary-flat.pdf](https://www.nature.com/documents/nr-reporting-summary-flat.pdf)

Life sciences study design

All studies must disclose on these points even when the disclosure is negative.

Sample size	A total of 13 mice and 1 zebrafish larva were used in this study to demonstrate the performance of the AO module described in this manuscript, as described in the following. 2P imaging data in the manuscript were acquired from 1 zebrafish larva and 2 mice. An additional zebrafish larva and mouse were used to further validate the reproducibility of our results (data not shown), for a total of 3 and 13 different fields-of-view for the zebrafish and mouse experiments, respectively. 3P cortical imaging data in the manuscript were acquired from 5 mice. An additional 3 mice were used to further validate the reproducibility of our results (data not shown), for a total of 32 different fields-of-view. Subcortical 3P imaging data in the manuscript were acquired from 3 mice. An additional mouse was used to further validate the reproducibility of our results (data not shown), for a total of 15 fields-of-view. Lastly, the 3P spinal cord imaging data in the manuscript were acquired from 2 mice for the structural imaging experiments, and 1 mouse for the calcium imaging experiments. An additional 2 mice was used to further validate the reproducibility of our results (data not shown), for a total of 10 fields-of-view. We did not perform any calculation to determine sample size. In some cases (e.g., 3P cortical imaging a depths ranging from 600-870 um) we used larger sample size (up to n=8 mice) with the purpose of showing the robustness in the performance of our module. The sample size used was sufficient to support the results in our study and further replication was not necessary.
Data exclusions	None
Replication	We used a total of 20 mice and 2 zebrafish larva. Using the AO module described in this manuscript, we consistently improved the image signal and contrast in these samples. For the purpose of this study, which aims to show the performance of the AO module to improve image quality in 2P and 3P fluorescence microscopy, rather than studying a specific biological question, the sample size was sufficient.
Randomization	Randomization does not apply to this study, since we did not seek to answer a specific biological question, but instead validate the ability of our technology to improve imaging performance in vivo.
Blinding	Blinding does not apply to this study since group allocation was not needed.

Reporting for specific materials, systems and methods

We require information from authors about some types of materials, experimental systems and methods used in many studies. Here, indicate whether each material, system or method listed is relevant to your study. If you are not sure if a list item applies to your research, read the appropriate section before selecting a response.

Materials & experimental systems

n/a	Involved in the study
<input checked="" type="checkbox"/>	<input type="checkbox"/> Antibodies
<input checked="" type="checkbox"/>	<input type="checkbox"/> Eukaryotic cell lines
<input checked="" type="checkbox"/>	<input type="checkbox"/> Palaeontology and archaeology
<input type="checkbox"/>	<input checked="" type="checkbox"/> Animals and other organisms
<input checked="" type="checkbox"/>	<input type="checkbox"/> Human research participants
<input checked="" type="checkbox"/>	<input type="checkbox"/> Clinical data
<input checked="" type="checkbox"/>	<input type="checkbox"/> Dual use research of concern

Methods

n/a	Involved in the study
<input checked="" type="checkbox"/>	<input type="checkbox"/> ChIP-seq
<input checked="" type="checkbox"/>	<input type="checkbox"/> Flow cytometry
<input checked="" type="checkbox"/>	<input type="checkbox"/> MRI-based neuroimaging

Animals and other organisms

Policy information about [studies involving animals](#); [ARRIVE guidelines](#) recommended for reporting animal research

Laboratory animals

Zebrafish (*Danio rerio*) larva, Tg(β -actin:HRAS-EGFP), 4-day old
The following mice were used (Jackson Laboratories): Thy1-YFP-H, Thy1-GFP-M, Gad2-IRES-Cre X Ai14, C57BL/6J; with ages ranging from 5 to > 30 weeks old. Mice were housed in cages in groups of 1–5 before surgery, under a normal light cycle, at a temperature of about 73 F and humidity of 37%. Both males and females were used, however, since the purpose of this study was to show the performance of an AO module to improve image quality rather than studying a specific biological question, we did not record the sex of each animal for every single experiment.

Wild animals

No wild animals were involved in this study.

Field-collected samples

No field-collected samples were involved in this study.

Ethics oversight

All animal experiments were conducted according to the National Institutes of Health guidelines for animal research. Procedures and protocols on mice and zebrafish were approved by the Institutional Animal Care and Use Committee at Janelia Research Campus, Howard Hughes Medical Institute; and the Animal Care and Use Committee at the University of California, Berkeley.

Note that full information on the approval of the study protocol must also be provided in the manuscript.